
Collagen Type II Cleavage
ITEM # 60-1017

INSTRUCTIONS FOR USE

IBEX C2C Human Urine Sandwich Assay, or IB-C2C-HUSA™, is a colour-monitoring immunoassay optimized for the quantitative determination of C2C neopeptide fragments present in human urine samples. This product is not suitable for testing serum. This procedure has been developed for *in-vitro* research use only and should be performed by trained professionals.

BACKGROUND AND PRINCIPLE OF THE ASSAY

Joint cartilage is composed of a type II collagen-based fibrillar network complexed to proteoglycans. Type II collagen consists of 3 identical α chains arranged in a triple helix that form fibrils. Within this fibrillar meshwork resides the large aggregating proteoglycan aggrecan. In arthritis, type II collagen is extensively cleaved and destroyed by the activity of collagenases, which results in loss of type II collagen (Billinghurst et al., 1997). Different collagenases, namely MMP-1, MMP-8 and MMP-13, can cleave type II collagen. Fragments containing the neopeptide created at the C terminus of the $\frac{3}{4}$ piece of type II collagen cleavage product (C2C neopeptide fragments) are recognized by IBEX C2C antibodies which are specific to type II collagen.

IB-C2C-HUSA™ is a colour-monitoring sandwich immunoassay that measures the C2C neopeptide fragments longer than 20 amino acids, present in human urine samples. It has been reported that a 45-mer peptide is the most abundant C2C neopeptide fragment in human urine samples (Nemirovskiy et al., 2007). The C2C levels in urine (called uC2C) vary with age. In normal healthy individuals, uC2C levels are very high in growing babies and children, but decrease substantially in healthy adults. Very active individuals may also have increased uC2C levels. Nevertheless, there is a strong correlation of increased uC2C levels with increased risks of having Osteoarthritis (OA) and/or Rheumatoid Arthritis (RA), despite the predictable profile of uC2C with age (Cibere et al., 2006; Cibere et al., 2009; Nemirovskiy et al., 2007 and unpublished data).

C2C biomarker may be used to analyze degradation in articular cartilages (Dejica et al., 2008). IB-C2C-HUSA™ alone or in combination with C1,2C and CPII assays may provide additional important data to clinical information or drug development.

IB-C2C-HUSA™ is based on the direct or 'antigen sandwich' principle (Figure 1). The assay uses a 96-well microplate (12 removable strips of 8 wells). The solid phase is coated with a cocktail of antibodies that effectively capture the C2C fragments. The Tracer antibody (Ab conjugated to HRP) provides further specificity to the assay.

Out-of-plate pre-dilution method is recommended for this assay. In a mixing plate, the orange-coded Matrix buffer (30 μ L) is first added to each well. Upon addition of standards or samples (30 μ L), the orange colour will change visually, confirming the addition and thus minimizing human errors due to repeated pipetting. Dilution buffer (60 μ L) is then added to each well. After gentle mixing, the pre-mix solutions are transferred to the coated microplate (100 μ L /well). An incubation step [1 hour at room temperature (RT°C), no shaking] allows the specific capture of C2C fragments to the solid phase. After removing unwanted residues through washing, Tracer Ab-HRP is added (100 μ L /well) to specifically bind to any C2C complexes captured to the solid phase. After removing unbound tracer, TMB substrate solution (100 μ L /well) is then added to reveal colour intensity proportional to the amount of the C2C fragments present in the test sample. Based on the values obtained with the standard set included in the assay, the content of C2C fragments may be estimated.




Figure 1: Principle of C2C sandwich assay

IBEX Pharmaceuticals Inc.

MATERIALS

Specimen collection

Second morning void urine samples is recommended. The urine samples should be stored at 4°C for less than 24 hours but they should be stored frozen (-20°C) for longer storage terms. Prior to use, urine samples should be thawed gradually to room temperature; the samples should be shaken and sedimentation allowed for a minimum of 30 minutes or centrifugation at 500 g for 5 minutes.

Supplied Reagents

Kit components	Format	Quantity
Coated microplate	Ready-to-use	12 strips of 8 wells (96 wells)
Matrix buffer	Ready-to-use	min. 3.5 mL
Standard set	Ready-to-use	7 x min. 130 µL/vial
Positive control - 1	Ready-to-use	min. 130 µL/vial
Positive control - 2	Ready-to-use	min. 130 µL/vial
Dilution buffer	Ready-to-use	min. 20 mL
Tracer Ab-HRP	Concentrated	min. 310 µL
TMB	Ready-to-use	min. 14 mL
Stop solution	Ready-to-use	min. 14 mL
Wash buffer (50x concentrate)	Concentrated	2 x min. 25 mL
Mixing plate	Ready-to-use	1 x 96 wells

Description

- Coated microplate: 12 strips of 8 breakable wells coated with a cocktail of antibodies. The coated microplate has been sealed into Zipper aluminum pouch with desiccant. Tightly close unused strips in the zipper pouch with desiccant to prevent moisture exposure and store at 2-8°C.
- Matrix buffer: 1 vial (min. 3.5 mL) ready-to-use and orange colour-coded reagent to be added to each well.
- Standard set: A set of seven (7) vials (min. 130 µL/each) of ready-to-use, blue colour-coded standards with known concentration of C2C peptide fragments:

Standard 1	=	5000 pg/mL
Standard 2	=	2500 pg/mL
Standard 3	=	1250 pg/mL
Standard 4	=	625 pg/mL
Standard 5	=	312 pg/mL
Standard 6	=	156 pg/mL
Standard 7	=	0 pg/mL

Store any unused portion at 2-8°C.

IBEX Pharmaceuticals Inc.

- Positive control - 1 (PC-1): 1 vial (min. 130 µL) ready-to-use High PC. Store any unused portion at 2-8°C.
- Positive control - 2 (PC-2): 1 vial (min. 130 µL) ready-to-use Low PC. Store any unused portion at 2-8°C.
- Dilution buffer: 1 vial (min. 20 mL) blue colour-coded ready-to-use Dilution buffer. Used as an assay buffer to be added to the Mixing plate; also used as an HRP diluent for dilution of concentrated Tracer Ab-HRP. This Dilution buffer may be used for pre-dilution of samples, should this be necessary. Store any unused portion at 2-8°C.
- Tracer Ab-HRP: 1 vial (min. 310 µL) concentrated Tracer Ab-HRP conjugate to be diluted 1:40 with Dilution buffer. For example, mix 300 µL concentrated Tracer Ab-HRP with 12 mL Dilution buffer. Store any unused concentrated Tracer at 2-8°C.
- TMB: 1 vial (min. 14 mL) ready-to-use HRP-substrate solution containing Tetra methylbenzidine (TMB) in a proprietary buffer. Keep away from direct sunlight. Store any unused portion at 2-8°C.
- Stop solution: 1 vial (min. 14 mL) ready-to-use Stop solution (diluted sulphuric acid). Handle all acid solutions with caution. Store any unused portion at 2-8°C.
- Wash buffer (50x concentrate): 2 vials (min. 25 mL/each) concentrated Wash solution containing buffered saline with a non-ionic detergent and a preservative.
- Mixing plate: 1 polypropylene non coated 96 well round bottom mixing plate.

Required Equipment not supplied

- Pipette capable of accurately dispensing 30 µL
- Multichannel pipettes capable of accurately dispensing 30-100 µL and 350 µL
- Graduated cylinders for reconstituting or diluting reagents
- Distilled or deionized water
- Vortex mixer
- Automatic washer or equivalent plate washing system
- ELISA plate reader suitable for 96-well plate formats and capable of measuring at 450 nm and 650 nm.
- 2-8°C environment(s) for product storage

WARNINGS AND PRECAUTIONS

This kit is intended for *in-vitro* research use by trained professionals only.

Storage

Store all kit reagents at 2-8°C. Do not freeze. Store any unused portion at 2-8°C. Seal the pouch containing unused coated wells (with desiccant) carefully. Close all vials well.

Kit stability

The kit is currently granted a 6-month shelf life. Once opened and stored appropriately, the kit reagents are stable for 1 month or up to the expiry date stated on the box, whichever comes first.

IBEX Pharmaceuticals Inc.

Frozen kit

While the kit should not be frozen, a kit that has been accidentally frozen once still performs as usual.

ASSAY PROCEDURE

The following procedure must be followed in order to minimize errors / variation and to optimize reliable results.

Reagent preparation:

1. Bring all kit components and samples to room temperature (18 – 25°C) before use for at least 30 minutes.
2. Wash solution: Dilute the Wash buffer (50x) 1:49 (v/v) with deionized water. For example, in a graduated cylinder, mix deionized water with 20 mL Wash buffer (50x) to a final volume of 1000 mL. Crystallization may occur at the bottom of the 50X concentrate vials. Prior to its use, simply dissolve the crystals by immersing the vials in warm water for a few minutes and mix well. Unused portion should be stored at 2-8°C and may be used within one week.
Note: It is important to always follow the same washing method (manual or automated) in order to minimize variation in OD signals.
3. Coated microplate: Remove required strips from pouch. Return unused strips and desiccant to original pouch. Seal pouch carefully after each use. Store at 2-8°C.
4. Standard set: All standards should be assayed in duplicate.
5. Urine samples: We recommend testing all samples in duplicate. Occasionally, some samples may require additional dilution prior to testing, in which case, optimal dilution factor for such samples has to be pre-determined.
6. Tracer Ab-HRP: Prepare the Tracer Ab-HRP working solution before use by diluting the Tracer Ab-HRP concentrate 1:40 with Dilution buffer. For example: add 300 µL to 12 mL of Dilution buffer; this is sufficient for testing 1 full microplate.

Test procedure:

1. Recommended positions of C2C standards and PC on Coated microplate:

	1	2	3
A	Std 1	Std 1	PC-2
B	Std 2	Std 2	PC-2
C	Std 3	Std 3	
D	Std 4	Std 4	
E	Std 5	Std 5	
F	Std 6	Std 6	
G	Std 7	Std 7	
H	PC-1	PC-1	


2. Dispense 30 µL of orange colour-coded Matrix buffer into each well of the Mixing plate. A multi-channel pipette may be used.
3. Dispense 30 µL of each Standard and Positive control into respective wells, using a separate pipette tip for each addition (including replicates). Dispense 30 µL of each test sample into individual wells beginning with well C3, using a separate pipette tip for each addition. Rinse tips by gently pipetting up

IBEX Pharmaceuticals Inc.

and down at least 3 times. Orange colour will change confirming the addition of the Standard, PC or sample. Different samples may show different colour change.

4. Dispense 60 µL of Dilution buffer into each well. A multi-channel pipette may be used.
5. Mix by gently tapping the side of the Mixing plate.
6. Using a multi-channel pipette, transfer 100 µL of the mixture from each well of the Mixing plate to the corresponding wells of the Coated microplate. Change pipette tips after each transfer.
7. Incubate at room temperature (18-25°C) for 1 hour (without agitation).
8. Aspirate and wash 5 times with 350 µL/well diluted Wash buffer (1x). Automated washers should be adjusted to fill each well completely without overflowing and with soaking time set to zero (0). Wash instrument and procedure should be verified for IB-C2C-HUSA™ prior to any study.
9. Add 100 µL of diluted Tracer Ab-HRP to each well.
10. Incubate at room temperature (18-25°C) for 30 minutes (without agitation).
11. Aspirate and wash 5 times with 350 µL/well diluted Wash buffer (1x). Automated washers should be adjusted to fill each well completely without overflowing and with soaking time set to zero (0). Wash instrument and procedure should be verified for IB-C2C-HUSA™ prior to any study.
12. After the final wash, gently tap the microplate upside down on absorbent paper to remove any remaining liquid, taking care not to dislodge strips from the holder.
13. Dispense 100 µL of TMB into each well. Protect the microplate from direct strong light sources to avoid an increased background signal (undesired oxidation of TMB).
14. Incubate the microplate at room temperature (18-25°C) for 30 minutes.
15. Stop the reaction by dispensing 100 µL of Stop solution into each well.
16. After adding the Stop solution, read the absorbance on a microplate reader at 450 nm and optionally at 650 nm for background value due to the microplate alone. Use the net absorbance values (OD 450 nm – OD 650 nm) for all further calculations. Absorbance must be read within 30 minutes for accurate measurement.

IBEX Pharmaceuticals Inc.

SUMMARY OF PROCEDURE (Please read the whole instruction carefully for more details)**IBEX Pharmaceuticals Inc.**

RESULTS

Calculation of results

A four- or five-parameter logistic curve fit (4PL or 5PL) can be used. Alternatively, plot the mean absorbance values of the standards (Y axis) against the corresponding C2C concentrations (X axis) and determine the C2C concentration of the controls and samples by interpolation. This is the C2C concentration (pg/mL) in each reaction well. Multiply this value with the dilution factor to obtain the C2C concentration present in the samples/controls.




Figure 1: Typical Standard Curve for IB-C2C-HUSA™ Assay
(Not for use in calculation of actual results)

TEST PERFORMANCE

Detection limit

The lower limit of quantification is 70 pg/mL

Specificity

No cross reactivity with corresponding neoepitope peptides derived from collagen I and collagen III is detected.

IBEX Pharmaceuticals Inc.

Spiking recovery

Human urine samples were spiked with different known amounts of C2C and assayed.

Urine sample	Observed [pg/mL]	Expected [pg/mL]	Recovery (%)
1	467	-	-
	4184	4217	99.2
	2363	2467	95.7
2	1704	-	-
	5284	5454	96.8
	3717	3704	100.3

Linearity dilution

Urine samples were diluted with Dilution buffer and assayed.

Sample ID	Dilution Factor	net OD 450-650nm	Conc [pg/mL]	Corrected conc [pg/mL]
h-u001	neat	0.518	1954	1954
	2	0.262	1021	2041
	4	0.128	535	2138
h-u002	neat	0.643	2408	2408
	2	0.309	1192	2383
	4	0.141	580	2320
	8	0.08	308	2464
h-u003	neat*	1.004	3722	3722
	2	0.909	3378	6756
	4	0.526	1982	7926
	8	0.236	928	7422

*: This sample exhibits a hook effect at neat concentration. Additional dilution corrects the problem.

Precision**Intra-lot**

Sample	Mean (pg/mL)	SD (pg/mL)	CV (%)
1	4931	16	< 1.0
2	2771	63	2.3
3	1093	39	3.6
4	592	22	3.7

Inter-lot

Sample	Mean (pg/mL)	SD (pg/mL)	CV (%)
1	4922	20	< 1.0
2	2732	81	3.0
3	1110	41	3.7
4	612	39	6.4

IBEX Pharmaceuticals Inc.

REFERENCES

- Bermingham J.D., Vilim V. and Kraus V. (2006) Collagen Biomarkers for Arthritis Applications. *Biomarker Insights* 1:61-76.
- Billinghurst R.C., Dahlberg I., Ionescu M., Reiner A., C Bourne R., Rorabeck C., Mitchell P., Hambor J., Diekmann O., Tschesche H., Chen J., Van Wart H. and Poole A.R. (1997). Enhanced cleavage of type II collagen by collagenases in osteoarthritic articular cartilage. *J. Clin. Invest.* 99: 1535-1545.
- Cibere J., Zhang H., Thorne A., Garnero P., Poole A.R., Lobanok T., Kopec J.A., Singer J., Wong H., Guermazi A., Peterfy C., Nicolaou S., Munk P., and Esdaile J.M. (2006). Urine CTX-II and Urine C2C are Differentially Associated with Radiographic and Pre-Radiographic Knee Osteoarthritis (OA): Results of a Population-Based Study Using MRI. *Arthritis and Rheumatism*: 54(9S): 265
- Cibere J., Zhang H., Garnero P., Poole A.R., Lobanok T., King L., Saxne T., Kraus V.B., Way A., Thorne A., Wong H., Singer J., Kopec J., Guermazi A., Peterfy C., Nicolaou C., Munk P., and Esdaile J.M. (2009). Association of biomarkers with pre-radiographically defined and radiographically defined knee osteoarthritis in a population-based study. *Arthritis and Rheumatism* 60: 1372-1380.
- Dejica V.M., Mort J.S., Laverty S., Percival M.D., Antoniou J., Zukor D.J. and Poole A.R. (2008). Cleavage of type II collagen by cathepsin K in human osteoarthritic cartilage. *Am. J. Pathol.* 173: 161-9.
- Kojima, T., Mwale F., Yasuda T., Girard C., Poole A.R. and Laverty S. (2001). Early degradation of type IX and type II collagens with onset of experimental inflammatory arthritis. *Arthritis Rheum.* 44: 120-127.
- Mullan R.H., Matthews C., Bresnihan B., FitzGerald O., King L., Poole A.R., Fearon U. and Veale D.J. (2007). Early changes in serum type II collagen biomarkers predict radiographic progression at one year in inflammatory arthritis patients after biologic therapy. *Arthritis and Rheumatism* 56: 2919-28.
- Nemirovskiy O.V., Dufield D.R., Sunyer T., Aggarwal P., Welsch D.J. and Mathews W.R. (2007). Discovery and development of a type II collagen neopeptide (TIINE) biomarker for matrix metalloproteinase activity: from in vitro to in vivo. *Anal Biochem.* 361(1): 93-101.
- Poole A.R., Alini M., and Hollander A.H. (1995). Cellular biology of cartilage degradation. In: Henderson B., Edwards, J.C.W., Pettipher, E.R., (Eds.). *Mechanisms and models in rheumatoid arthritis*. London: Academic Press; pp. 163-204.
- Song X-Y., Zeng L., Jin W., Thompson J., Mizel D.E., Lei K-J., Billinghurst R.C., Poole A.R. and Wahl S.M. (1999). Secretory leukocyte protease inhibitor suppresses the inflammation and joint damage of bacterial cell wall-induced arthritis. *J. Exp. Med.* 190: 535-542.
- Verstappen S.M.M., Poole A.R., Ionescu M., King L.E., Abrahamowicz M., Hofman D.M., Bijlsma J.W.J., Labeber F.P, J.G and the Utrecht Rheumatoid Arthritis Cohort Study Group (2006). Radiographic joint damage in rheumatoid arthritis is associated with differences in cartilage turnover and can be predicted by serum biomarkers: an evaluation from 1 to 4 years after diagnosis. *Arthritis Research and Therapy*: R31.

IBEX Pharmaceuticals Inc.